



Member State
Denmark

OIML Certificate N°
R76/1992-DK-02.02

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **Danish Agency for Trade and Industry**
Address: Dahlerups Pakhus
Langelinie Allè 17
DK-2100 Copenhagen Ø
Denmark

Person responsible: P. Claudi Johansen

Applicant

Name: **ESIT Electronics**
Address: Mühürdar Cad. No.91
Kadiköy, Istanbul
Turkey

Manufacturer

of the certified pattern: **ESIT Electronics**

Identification

of the certified pattern: **Nonautomatýc weighing instrument**
Type: PWI
Further characterisites are set out on page 2.

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the assoiated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 1992, including amendment 1 (1994)
for accuracy class III and IIII

This certificate relates only to the metrological and technical characterisites of the pattern of the instrumentation concerned, as covered by the relevant OIML International Recommendation.
This certificate does not bestow any form of legal international approval.

**OIML Certificate N°
R76/1992-DK-02.02**

The conformity was established by tests described in the associated test report from DELTA, DK N° DANAK-195482, that includes 10 pages and 1 annex (61 pages).

The issuing authority


Danish Agency for Trade and Industry

Date: 18. APR. 2002

The OIML member

P. Claudi Johansen

Date: 18. APR. 2002



Characteristics:

Accuracy class		III
Weighing range		Single-interval or multi-interval
Number of intervals	Max /e	≤ 6000 (class III), 1000 (class IIII)
Minimum input-voltage	ΔU_{\min}	1 μ V per verification scale interval
Measuring range	Min./Max.	0.5 mV / 100mV
Excitation voltage	U_{exe}	10 Vdc
Load cell impedance	Min./Max.	35 ohm / 1000 ohm
Load cell connecting system		4-wires or 6 wires system, screened
Module fractional factor	P_1	0.5 for the indicator
Interface		Protective, according to paragraph 5.3.6
Connected load cells		Shall comply with R60
Temperature range for the indicator		-10 °C / +40°C

Important note:

Apart from the mention of certificate's reference number and the name of the OIML. Member State in which the certificate was issued, partial quotation of the certificate or the associated test report is not permitted, though they may be reproduced in full.